

土城院區與林口院區一致性(Correlation):

1. 陰性檢體一致性達 100%

林口 8800 result		土城 6800 result	
Sample ID	Result	Sample ID	Result
24	HBV001	HBV001	Target Not Detected
25	HBV002	HBV002	Target Not Detected
26	HBV003	HBV003	Target Not Detected
27	HBV005	HBV005	Target Not Detected
28	HBV006	HBV006	Target Not Detected
29	HBV007	HBV007	Target Not Detected
30	HBV008	HBV008	Target Not Detected
31	HBV009	HBV009	Target Not Detected
32	HBV010	HBV010	Target Not Detected
33	NEG000	NEG000	Target Not Detected

2. 陽性檢體 log10 值 $R^2 > 0.95$ 無明顯差異

